

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-09032024-013

Gage ID PEA-2419001778 Gage S/N 2419001778 Description UV Light Meter Operating Procedure: OP-119 Unit of Meas. Microwatts/cm2 Manufacturer General Cal. Date 9/3/2024 Next Due 3/3/2025 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Calibrated black light serial no. 10960933 was also used during this calibration.
Calibration due date for black light is on 06/17/2025

Black light is traceable to NIST by the following numbers: 258648, 258649, 258650

Standard ID 12"	Uncertainty	Units	Type V
Limited Use? No	Minimum 4659	Nominal 4904	Maximum 5149
Ref Type Long Wave	As Found 5013	Accuracy 109	Fail Before No
	As Left 5013	Accuracy 109	Fail After No
Gage ID of Standard REF-1944278			
Std Due Date 6/17/2025	Gage S/N 1944278/1944 280	NIST No. 258648, 258649, 258650	

Standard ID 15"	Uncertainty	Units	Type V
Limited Use? No	Minimum 3142	Nominal 3307	Maximum 3472
Ref Type Long Wave	As Found 3401	Accuracy 94	Fail Before No
	As Left 3401	Accuracy 94	Fail After No
Gage ID of Standard REF-1944278			
Std Due Date 6/17/2025	Gage S/N 1944278/1944 280	NIST No. 258648, 258649, 258650	

Calibrated By Taichi Daimo

Signature 

Date: 9/3/2024

APPROVED
By Kayla Myers at 10:57 am, Sep 06, 2024