TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-07262024-004

Gage ID PEA-CES-105643

Gage S/N CES-105643

Description 10 1/2 In. DC Magnetic Coil

Operating OP-116 Procedure:

Unit of Meas. Amps

Manufacturer CES-Cut-N-Edge Service

Cal. Date 7/26/2024 Next Due 1/26/2025

Cal. Freq. 6.00

Months

Location Lab

Environmental Conditions

Temperature

68 +/- 2 deg F

Humidity

20-55%

Approved Yes

Customer Info. Peak NDT

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13137 JOB 13701

AS FOUND: PASSED

AS LEFT: PASSED

Mag coil putting out magnetic field from both the pos. and neg. sides of coil equal to or greater than 50 gauss for calibration.

Calibrated 50-0-50 gauss meter serial no. 09-1234 was used for this calibration.

Calibration due date for gauss meter is on 05/10/25

Gauss meter is traceable to NIST by the following number: 101912

Calibrated By Taichi Daimo Signature

7/26/2024

APPROVED

By Kayla Myers at 2:35 pm, Jul 29, 2024