

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-08302023-005

<p>Gage ID PEA-J02108 Gage S/N J02108 Description 5 Step Block Operating Procedure: OP-126 Unit of Meas. Inch Manufacturer N/A Cal. Date 8/30/2023 Next Due 8/30/2024 Cal. Freq. 1.00 Years Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 12069 JOB 11683

Standard ID	0.100"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.098	Nominal	0.100	Maximum	0.102
Ref Type	6" Digital Caliper	As Found	0.100	Accuracy	0.000	Fail Before	No
		As Left	0.100	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639
Std Due Date 10/7/2023 **Gage S/N** 08341639 **NIST No.** CK303-38600-320

Standard ID	0.200"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.198	Nominal	0.200	Maximum	0.202
Ref Type	6" Digital Caliper	As Found	0.200	Accuracy	0.000	Fail Before	No
		As Left	0.200	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639
Std Due Date 10/7/2023 **Gage S/N** 08341639 **NIST No.** CK303-38600-320

Standard ID	0.300"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.298	Nominal	0.300	Maximum	0.302

Ref Type	6" Digital Caliper	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/7/2023 Gage S/N 08341639 NIST No. CK303-38600-320

Standard ID	0.400"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	6" Digital Caliper	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/7/2023 Gage S/N 08341639 NIST No. CK303-38600-320

Standard ID	0.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	6" Digital Caliper	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/7/2023 Gage S/N 08341639 NIST No. CK303-38600-320

Calibrated By Taichi Daimo Signature  Date: 8/30/2023

APPROVED
By Kayla Myers at 7:37 am, Aug 31, 2023