

### TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

## Certificate of Calibration

Certificate No TSP-07262024-005

<p><b>Gage ID</b> PEA-20847  <b>Gage S/N</b> 20847  <b>Description</b> 10 lb. Test Weight  <b>Operating Procedure:</b> OP-118  <b>Unit of Meas.</b> lbs  <b>Manufacturer</b> Parker  <b>Cal. Date</b> 7/26/2024  <b>Next Due</b> 7/26/2027  <b>Cal. Freq.</b> 3.00      Years  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b></p> <p>Temperature    68 +/- 2 deg F  Humidity        20-55%</p> <p style="text-align: center;"><b>Approved</b>    Yes  <b>Customer Info.</b> Peak NDT</p>
--	---

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

SO 13137 JOB 13701

AS FOUND: PASSED      AS LEFT: PASSED

Using calibrated scale serial no. 7129341906  
Test weight as found 10.07 lbs.  
Test weight as left 10.07 lbs.

Calibration due date for scale is on 4/15/2025.  
Scale is traceable to NIST by the following numbers: LA22-258

Calibrated By Taichi Daimo      Signature       Date: 7/26/2024

**APPROVED**  
By Kayla Myers at 2:39 pm, Jul 29, 2024