

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-08302023-006

<p>Gage ID PEA-22254 Gage S/N 22254 Description UT Thickness Gage Operating Procedure: OP-121 Unit of Meas. inch Manufacturer Dakota Ultrasonics Cal. Date 8/30/2023 Next Due 2/29/2024 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 12069 JOB 11683

Standard ID	0.250"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.248	Nominal	0.250	Maximum	0.252
Ref Type	8 Step Block	As Found	0.250	Accuracy	0.000	Fail Before	No
		As Left	0.250	Accuracy	0.000	Fail After	No
Gage ID of Standard	Ref-419719						
Std Due Date	12/7/2023	Gage S/N	419719	NIST No.	115078		
Standard ID	0.750"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.748	Nominal	0.750	Maximum	0.752
Ref Type	8 Step Block	As Found	0.751	Accuracy	0.001	Fail Before	No
		As Left	0.751	Accuracy	0.001	Fail After	No
Gage ID of Standard	Ref-419719						
Std Due Date	12/7/2023	Gage S/N	419719	NIST No.	115078		
Standard ID	1.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.998	Nominal	1.000	Maximum	1.002

Ref Type	8 Step Block	As Found	1.001	Accuracy	0.001	Fail Before	No
		As Left	1.001	Accuracy	0.001	Fail After	No

Gage ID of Standard Ref-419719

Std Due Date 12/7/2023 Gage S/N 419719 NIST No. 115078

Standard ID	1.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	1.498	Nominal	1.500	Maximum	1.502
Ref Type	8 Step Block	As Found	1.500	Accuracy	0.000	Fail Before	No
		As Left	1.500	Accuracy	0.000	Fail After	No

Gage ID of Standard Ref-419719

Std Due Date 12/7/2023 Gage S/N 419719 NIST No. 115078

Standard ID	2.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	1.998	Nominal	2.000	Maximum	2.002
Ref Type	8 Step Block	As Found	2.000	Accuracy	0.000	Fail Before	No
		As Left	2.000	Accuracy	0.000	Fail After	No

Gage ID of Standard Ref-419719

Std Due Date 12/7/2023 Gage S/N 419719 NIST No. 115078

Calibrated By Taichi Daimo Signature  Date: 8/30/2023

APPROVED
By Kayla Myers at 7:38 am, Aug 31, 2023