

### TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

## Certificate of Calibration

Certificate No TSP-08012024-003

<p><b>Gage ID</b> PEA-210701049  <b>Gage S/N</b> 210701049  <b>Description</b> Visible Light Meter  <b>Operating Procedure:</b> OP-130  <b>Unit of Meas.</b> Foot Candles  <b>Manufacturer</b> General  <b>Cal. Date</b> 8/1/2024  <b>Next Due</b> 2/1/2025  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>  Temperature 68 +/- 2 deg F  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT</p>
---	--

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

Using calibrated light meter serial no. 1944278 with sensor 1944280 the following readings were taken.

Readings As Found (foot candles)

	Master	Actual
High	96.7	94.4
Low	34.5	32.7

**APPROVED**  
*By Kayla Myers at 7:32 am, Aug 02, 2024*

Readings As Left (foot candles)

	Master	Actual
High	96.7	94.4
Low	34.5	32.7

Tolerance: +/- 5% or +/- 2 foot candles whichever is greater  
Cal due date for light meter is 06/17/2025  
NIST Traceable through : 258648, 258649, 258650

Calibrated By Taichi Daimo

Signature



Date: 8/1/2024