TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-01072025-004

Gage S/N 3006211618

Gage S/N 3006211618

Description 12" Digital Caliper

Operating OP-110 Procedure:

Unit of Meas. Inch Manufacturer Insize

Cal. Date 1/7/2025
Next Due 7/7/2025

Cal. Freq. 6.00 Months

Location Lab

Environmental Conditions

Temperature 68 +/- 2 deg F

Humidity 20-55%

Approved Yes
Customer Info. Peak NDT

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	1.000 "	Uncertainty			Units		Туре	V
Limited Use?	No	Minimum		0.999	Nominal	1.000	Maximum	1.001
Ref Type	GAGE BLOCKS	As Found		1.000	Accuracy	0.000	Fail Before	No
		As Left		1.000	Accuracy	0.000	Fail After	No
Gage ID of Star	ndard REF-GBS-061259							
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.		LMUA-131223 2018-3-20, AFS-276 2018-3-		
Standard ID	3.000 "	Uncertainty			Units		Туре	V
Limited Use?	No	Minimum		2.999	Nominal	3.000	Maximum	3.001
Ref Type	GAGE BLOCKS	As Found		3.000	Accuracy	0.000	Fail Before	No
		As Left		3.000	Accuracy	0.000	Fail After	No
Gage ID of Star	ndard REF-GBS-061259							
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.		LMUA-131223 2018-3-20, AFS-276 2018-3-2		
Standard ID	5.000 "	Uncertainty			Units		Туре	V
Limited Use?	No	Minimum		4.999	Nominal	5.000	Maximum	5.001

							, ago	2 01	_
Ref Type	GAGE BLOCKS	As Found		5.000	Accuracy	0.000	Fail Before	No	
		As Left		5.000	Accuracy	0.000	Fail After	No	
Gage ID of Sta	ndard REF-GBS-061259								
Std Due Date 12/3/2027		Gage S/N	061259	NIST No.		LMUA-131223 2018-3-20, AFS-276 2018-3-			
Standard ID	7.000 "	Uncertainty			Units		Туре	V	_
Limited Use?	No	Minimum		6.999	Nominal	7.000	Maximum	7.001	
Ref Type	GAGE BLOCKS	As Found		7.000	Accuracy	0.000	Fail Before	No	
		As Left		7.000	Accuracy	0.000	Fail After	No	
Gage ID of Sta	ndard REF-GBS-061259								
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.		LMUA-131223 2018-3-20, AFS-276 2018-3			18-3-20
Standard ID	9.000 "	Uncertainty			Units		Туре	V	
Limited Use?	No	Minimum		8.999	Nominal	9.000	Maximum	9.001	
Ref Type	GAGE BLOCKS	As Found		9.000	Accuracy	0.000	Fail Before	No	
		As Left		9.000	Accuracy	0.000	Fail After	No	
Gage ID of Sta	ndard REF-GBS-061259								

Calibrated By Taichi Daimo

Std Due Date

Gage ID: PEA-3006211618

12/3/2027

Signature

NIST No.

Date:

LMUA-131223 2018-3-20, AFS-276 2018-3-20

1/7/2025

2 of 2

Page

APPROVED

By Kayla Myers at 1:14 pm, Jan 08, 2025

Gage S/N 061259