

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-01072025-002

<p><b>Gage ID</b> PEA-05-1556  <b>Gage S/N</b> 05-1556  <b>Description</b> 10-0-10 Gauss Meter  <b>Operating Procedure:</b> OP-117  <b>Unit of Meas.</b> Gauss  <b>Manufacturer</b> Magnaflux  <b>Cal. Date</b> 1/7/2025  <b>Next Due</b> 7/7/2025  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT</p>
--	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

<b>Standard ID</b> +5	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b>	4 <b>Nominal</b>	5 <b>Maximum</b> 6
<b>Ref Type</b> Gauss Meter	<b>As Found</b>	5 <b>Accuracy</b>	0 <b>Fail Before</b> No
	<b>As Left</b>	5 <b>Accuracy</b>	0 <b>Fail After</b> No
<b>Gage ID of Standard</b> Ref-09-1229	<b>Gage S/N</b> 09-1229	<b>NIST No.</b> 114537	
<b>Std Due Date</b> 10/10/2025			

<b>Standard ID</b> +10	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b>	9 <b>Nominal</b>	10 <b>Maximum</b> 11
<b>Ref Type</b> Gauss Meter	<b>As Found</b>	10 <b>Accuracy</b>	0 <b>Fail Before</b> No
	<b>As Left</b>	10 <b>Accuracy</b>	0 <b>Fail After</b> No
<b>Gage ID of Standard</b> Ref-09-1229	<b>Gage S/N</b> 09-1229	<b>NIST No.</b> 114537	
<b>Std Due Date</b> 10/10/2025			

<b>Standard ID</b> -5	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b>	4 <b>Nominal</b>	5 <b>Maximum</b> 6

Ref Type	Gauss Meter	As Found	5	Accuracy	0	Fail Before	No
		As Left	5	Accuracy	0	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/10/2025

Gage S/N 09-1229

NIST No. 114537

Standard ID	-10	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	9	Nominal	10	Maximum	11
Ref Type	Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/10/2025

Gage S/N 09-1229

NIST No. 114537

Calibrated By Taichi Daimo Signature  Date: 1/7/2025

**APPROVED**  
*By Kayla Myers at 1:13 pm, Jan 08, 2025*