

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-01102025-000

<p>Gage ID PEA-JZ02JF0034 Gage S/N JZ02JF0034 Description Pit Gauge Operating Procedure: OP-122 Unit of Meas. inch Manufacturer Gagemaker Cal. Date 1/10/2025 Next Due 1/10/2026 Cal. Freq. 1.00 Years Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	.10000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.09900	Nominal	0.10000	Maximum	0.10100
Ref Type	Mic Trac	As Found	0.10003	Accuracy	0.00003	Fail Before	No
		As Left	0.10003	Accuracy	0.00003	Fail After	No
Gage ID of Standard	REF-G141308						
Std Due Date	9/30/2025	Gage S/N	G141308	NIST No.	747V66-240502-00		

Standard ID	.20000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.19900	Nominal	0.20000	Maximum	0.20100
Ref Type	Mic Trac	As Found	0.20008	Accuracy	0.00008	Fail Before	No
		As Left	0.20008	Accuracy	0.00008	Fail After	No
Gage ID of Standard	REF-G141308						
Std Due Date	9/30/2025	Gage S/N	G141308	NIST No.	747V66-240502-00		

Standard ID	.30000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.29900	Nominal	0.30000	Maximum	0.30100

Ref Type	Mic Trac	As Found	0.30008	Accuracy	0.00008	Fail Before	No
		As Left	0.30008	Accuracy	0.00008	Fail After	No

Gage ID of Standard REF-G141308
 Std Due Date 9/30/2025

Gage S/N G141308 NIST No. 747V66-240502-00

Standard ID	.40000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.39900	Nominal	0.40000	Maximum	0.40100
Ref Type	Mic Trac	As Found	0.40003	Accuracy	0.00003	Fail Before	No
		As Left	0.40003	Accuracy	0.00003	Fail After	No

Gage ID of Standard REF-G141308
 Std Due Date 9/30/2025

Gage S/N G141308 NIST No. 747V66-240502-00

Standard ID	.50000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.49900	Nominal	0.50000	Maximum	0.50100
Ref Type	Mic Trac	As Found	0.50001	Accuracy	0.00001	Fail Before	No
		As Left	0.50001	Accuracy	0.00001	Fail After	No

Gage ID of Standard REF-G141308
 Std Due Date 9/30/2025

Gage S/N G141308 NIST No. 747V66-240502-00

Calibrated By Taichi Daimo Signature  Date: 1/10/2025

APPROVED
 By Kayla Myers at 11:10 am, Jan 10, 2025