

### TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

## Certificate of Calibration

Certificate No TSP-07112024-012

<p><b>Gage ID</b> PEA-339814  <b>Gage S/N</b> 339814  <b>Description</b> 4 Step Block  <b>Operating Procedure:</b> OP-126  <b>Unit of Meas.</b> inch  <b>Manufacturer</b> N/A  <b>Cal. Date</b> 7/11/2024  <b>Next Due</b> 7/11/2025  <b>Cal. Freq.</b> 1.00      Years  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT</p>
--	--

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

SO 13083 JOB 13644

<b>Standard ID</b>	.250 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.248	<b>Nominal</b>	0.250	<b>Maximum</b>	0.252
<b>Ref Type</b>	6" Digital Caliper	<b>As Found</b>	0.250	<b>Accuracy</b>	0.000	<b>Fail Before</b>	No
		<b>As Left</b>	0.250	<b>Accuracy</b>	0.000	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-08341639						
<b>Std Due Date</b>	10/10/2024	<b>Gage S/N</b>	08341639	<b>NIST No.</b>	CK303-38600-320		

<b>Standard ID</b>	.500 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.498	<b>Nominal</b>	0.500	<b>Maximum</b>	0.502
<b>Ref Type</b>	6" Digital Caliper	<b>As Found</b>	0.500	<b>Accuracy</b>	0.000	<b>Fail Before</b>	No
		<b>As Left</b>	0.500	<b>Accuracy</b>	0.000	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-08341639						
<b>Std Due Date</b>	10/10/2024	<b>Gage S/N</b>	08341639	<b>NIST No.</b>	CK303-38600-320		

<b>Standard ID</b>	.750 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.748	<b>Nominal</b>	0.750	<b>Maximum</b>	0.752

Ref Type	6" Digital Caliper	As Found	0.750	Accuracy	0.000	Fail Before	No
		As Left	0.750	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/10/2024

Gage S/N 08341639

NIST No.

CK303-38600-320

Standard ID	1.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.998	Nominal	1.000	Maximum	1.002
Ref Type	6" Digital Caliper	As Found	1.000	Accuracy	0.000	Fail Before	No
		As Left	1.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/10/2024

Gage S/N 08341639

NIST No.

CK303-38600-320

Calibrated By Taichi Daimo

Signature



Date: 7/11/2024

**APPROVED**

*By Kayla Myers at 3:07 pm, Jul 11, 2024*