



CALIBRATION CERTIFICATION

INSTRUMENT: MIZ-21C EDDY CURRENT TESTER

SERIAL NUMBER: 794722

CONDITION CODE: B

The calibration of this instrument has been performed using a documented, controlled Zetec procedure which meets or exceeds the manufacturer's requirements.

CUSTOMER: PEAK NDT
4407 LA-14 New Iberia, LA 70560

Zetec Calibration Laboratories have been audited and found in compliance with ISO/IEC 17025:2017. Accredited calibrations performed within the Lab's Scope of Accreditation are indicated by the presence of the Accrediting Body's name and Zetec's Certification Number on this Certificate of Calibration. Zetec hereby certifies that the equipment calibrated per this certificate meets or exceeds all manufacture specifications and has been calibrated using a documented, controlled Zetec procedure accomplished in compliance with the current Zetec Quality Program. This calibration has been performed using measurement standards whose accuracies are traceable to the International System of Units. The certificate results relate only to this calibrated instrument. There is no implied warranty that the instrument will maintain its specified tolerances during the calibration interval due to possible drift, environment, or other factors beyond Zetec control. Unless otherwise specified, Zetec utilizes the simple decision rule based on tolerance boundaries. All purchase order requirements have been met. This certificate may not be reproduced or distributed except in whole without the written authorization from Zetec.

N. I. S. T TRACEABILITY:

Standard Instrument	Serial Number	Calibration Date	Expiration
AMS062 TEMP/HUMIDITY LOGGER	140202359	04-25-2024	04-25-2025
AMS101 FUNCTION GENERATOR	MY53301138	08-23-2024	08-23-2025
4917 DIGITAL MULTIMETER	6392018	04-30-2024	04-30-2025

CALIBRATION DATE: 14 January 2025

CALIBRATED BY: GREG CRUMPACKER/SR. CALIBRATION TECH

EXPIRATION DATE: 14 January 2026

REVIEWED BY: JAMES LARSEN/SR. CALIBRATION TECH

Procedure#: SLP-3114-PR-Z Rev. C **WO#:** 23921

COMMENTS:

TEMPERATURE: 21.0 DEG C HUMIDITY: 31.9% RH
DIGITAL READOUT IN 0.9 TO 110% IACS (0.5 TO 70 MS/M), ACCURACY WITHIN +/-0.5% IACS FROM 0.9% TO 65% IACS AND WITHIN +/-1.0% OF VALUES OVER 65%
MIZ-21C SINGLE FREQUENCY SEALED
SOFTWARE VERSION: 1.4.1.24
FFU: 1.2.0.19
FPGA: 1.50

Calibration Certification Number: 8772

8226 Bracken Place SE | Suite 100
Snoqualmie WA 98065
Phone: 425.974.2700
Fax: 425.974.2701