

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-06032024-011

| | |
|---|---|
| Gage ID PEA-220942452 Gage S/N 220942452 Description 0-1" Groove Micrometer Operating Procedure: OP-111 Unit of Meas. inch Manufacturer insize Cal. Date 6/3/2024 Next Due 12/3/2024 Cal. Freq. 6.00 Months Location Lab | Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions |
|---|---|

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

0-1" Groove Micrometer Measured Values As Found

| Master | Actual | Deviation |
|--------|--------|-----------|
| +0.100 | +0.100 | +0.000 |
| +0.300 | +0.300 | +0.000 |
| +0.500 | +0.500 | +0.000 |
| +0.700 | +0.700 | +0.000 |
| +0.900 | +0.900 | +0.000 |


APPROVED
By Kayla Myers at 12:42 pm, Jun 04, 2024

Tol +0.001 / -0.001

NOTE: IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS.

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:

Reference Standard Serial #: 070111
GAGE BLOCKS
NIST Number: 81004
Cal Date: 1/26/2024 Due Date: 1/26/2027

Calibrated By Chad Adams Signature  Date: 6/3/2024