TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-06122024-000

Gage ID PEA-CES-105455 Gage S/N CES-105455

Description 10 1/2" Magnetic Coil

Operating Procedure:

OP-116

Unit of Meas. gauss

Manufacturer Cut-N-Edge

Cal. Date 6/12/2024 Next Due 12/12/2024

Cal. Freq. 6.00

Location Lab

Environmental Conditions

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 12984 Job 13188

AS FOUND: PASSED

AS LEFT: PASSED

Months

Mag coil putting out magnetic field from both the pos. and neg. sides of coil equal to or greater than 50 gauss for calibration.

Calibrated 50-0-50 gauss meter serial no. TSPSTD-01 was used for this calibration.

Calibration due date for gauss meter is on 07/26/2024

Gauss meter is traceable to NIST by the following number: Certification #138431

Calibrated By Chad Adams

Signature

Date:

6/12/2024

APPROVED

By Kayla Myers at 2:14 pm, Jul 30, 2024