

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-08122024-018

<p>Gage ID PEA-210701029 Gage S/N 210701029 Description White Light Meter Operating Procedure: OP-130 Unit of Meas. Foot Candles Manufacturer General Cal. Date 8/12/2024 Next Due 2/12/2025 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: right;">Approved Yes Customer Info. Peak NDT Solutions</p>
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13176 Job 13190

Using calibrated light meter serial no. 1705427 with sensor 1705430 the following readings were taken.

Readings As Found (foot candles)

	Master	Actual
High	81.3	80.6
Low	11.9	11.1

Readings As Left (foot candles)

	Master	Actual
High	81.3	80.6
Low	11.9	11.1

APPROVED
 By Kayla Myers at 8:37 am, Aug 13, 2024

Tolerance: +/- 5% or +/- 2 foot candles whichever is greater.

Calibration due date for light meter is on 8/8/2025
 Light Meter is traceable to NIST by the following numbers: Doc# 154789

Calibrated By Chad Adams Signature  Date: 8/12/2024