

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-02142025-015

<p>Gage ID PEA-21436 Gage S/N 21436 Description A/C Yoke Operating Procedure: OP-115 Unit of Meas. pounds Manufacturer Electro-Spect Cal. Date 2/14/2025 Next Due 8/14/2025 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13910 Job 13190

AS FOUND: PASSED AS LEFT: PASSED

A/C Yoke picked up and held required 10 lb. test weight for calibration.
Calibrated test weight serial no. 3452 was used for this calibration.
Test weight calibration is due on 8/8/2025
Test weight is traceable to NIST by the following numbers: Doc# 154781

Calibrated By Chad Adams Signature  Date: 2/14/2025

APPROVED
By Kayla Myers at 11:13 am, Mar 03, 2025