

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-02142025-009

| | |
|---|--|
| <p>Gage ID PEA-7632 Gage S/N 7632 Description 20-0-20 gauss meter Operating Procedure: OP-117 Unit of Meas. gauss Manufacturer R.B. Annis Cal. Date 2/14/2025 Next Due 8/14/2025 Cal. Freq. 6.00 Months Location Lab</p> | <p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p> |
|---|--|

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13910 Job 13190

| | | | | | | | |
|----------------------------|---------------------|--------------------|-----------|-----------------|--------|--------------------|----|
| Standard ID | +10 | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 9 | Nominal | 10 | Maximum | 11 |
| Ref Type | 20-0-20 Gauss Meter | As Found | 10 | Accuracy | 0 | Fail Before | No |
| | | As Left | 10 | Accuracy | 0 | Fail After | No |
| Gage ID of Standard | REF-TSPSTD-02 | | | | | | |
| Std Due Date | 8/8/2025 | Gage S/N | TSPSTD-02 | NIST No. | 154784 | | |

| | | | | | | | |
|----------------------------|---------------------|--------------------|-----------|-----------------|--------|--------------------|----|
| Standard ID | +20 | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 19 | Nominal | 20 | Maximum | 21 |
| Ref Type | 20-0-20 Gauss Meter | As Found | 20 | Accuracy | 0 | Fail Before | No |
| | | As Left | 20 | Accuracy | 0 | Fail After | No |
| Gage ID of Standard | REF-TSPSTD-02 | | | | | | |
| Std Due Date | 8/8/2025 | Gage S/N | TSPSTD-02 | NIST No. | 154784 | | |

| | | | | | | | |
|---------------------|-----|--------------------|---|----------------|----|----------------|----|
| Standard ID | -10 | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 9 | Nominal | 10 | Maximum | 11 |

| | | | | | | | |
|----------|---------------------|----------|----|----------|---|-------------|----|
| Ref Type | 20-0-20 Gauss Meter | As Found | 10 | Accuracy | 0 | Fail Before | No |
| | | As Left | 10 | Accuracy | 0 | Fail After | No |

Gage ID of Standard REF-TSPSTD-02

| | | | | | |
|--------------|----------|----------|-----------|----------|--------|
| Std Due Date | 8/8/2025 | Gage S/N | TSPSTD-02 | NIST No. | 154784 |
|--------------|----------|----------|-----------|----------|--------|

| | | | | | | | |
|--------------|---------------------|-------------|----|----------|----|-------------|----|
| Standard ID | -20 | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 19 | Nominal | 20 | Maximum | 21 |
| Ref Type | 20-0-20 Gauss Meter | As Found | 20 | Accuracy | 0 | Fail Before | No |
| | | As Left | 20 | Accuracy | 0 | Fail After | No |

Gage ID of Standard REF-TSPSTD-02

| | | | | | |
|--------------|----------|----------|-----------|----------|--------|
| Std Due Date | 8/8/2025 | Gage S/N | TSPSTD-02 | NIST No. | 154784 |
|--------------|----------|----------|-----------|----------|--------|

Calibrated By Chad Adams Signature  Date: 2/14/2025

APPROVED
 By Kayla Myers at 11:10 am, Mar 03, 2025