

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-02142025-016

<p>Gage ID PEA-0965-10 Gage S/N 0965-10 Description Magnetic Coil Operating Procedure: OP-116 Unit of Meas. gauss Manufacturer Drilco Cal. Date 2/14/2025 Next Due 8/14/2025 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13910 Job 13190

AS FOUND: PASSED AS LEFT: PASSED

Mag coil putting out magnetic field from both the pos. and neg. sides of coil equal to or greater than 50 gauss for calibration.

Calibrated 50-0-50 gauss meter serial no. TSPSTD-01 was used for this calibration.
Calibration due date for gauss meter is on 8/8/2025
Gauss meter is traceable to NIST by the following number: Certification #154783

Calibrated By Chad Adams Signature  Date: 2/14/2025

APPROVED
By Kayla Myers at 11:14 am, Mar 03, 2025