

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-02142025-010

<p>Gage ID PEA-4055 13 Gage S/N 4055 13 Description 5-step block Operating Procedure: OP-126 Unit of Meas. inch Manufacturer Curtis Industries Cal. Date 2/14/2025 Next Due 2/14/2026 Cal. Freq. 1.00 Years Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13910 Job 13190

Standard ID	.100 "	Uncertainty	
Limited Use?	No	Minimum	0.098
Ref Type	12" Digital Caliper	As Found	0.100
		As Left	0.100
Gage ID of Standard	REF-1095377	Units	Nominal
Std Due Date	8/8/2025		0.100
		Gage S/N	1095377
		NIST No.	154785
		Maximum	0.102
		Fail Before	No
		Fail After	No

Standard ID	.200 "	Uncertainty	
Limited Use?	No	Minimum	0.198
Ref Type	12" Digital Caliper	As Found	0.200
		As Left	0.200
Gage ID of Standard	REF-1095377	Units	Nominal
Std Due Date	8/8/2025		0.200
		Gage S/N	1095377
		NIST No.	154785
		Maximum	0.202
		Fail Before	No
		Fail After	No

Standard ID	.300 "	Uncertainty	
Limited Use?	No	Minimum	0.298
		Units	Nominal
			0.300
		Maximum	0.302

Ref Type	12" Digital Caliper	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date	8/8/2025	Gage S/N	1095377	NIST No.	154785
--------------	----------	----------	---------	----------	--------

Standard ID	.400 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	12" Digital Caliper	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date	8/8/2025	Gage S/N	1095377	NIST No.	154785
--------------	----------	----------	---------	----------	--------

Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	12" Digital Caliper	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date	8/8/2025	Gage S/N	1095377	NIST No.	154785
--------------	----------	----------	---------	----------	--------

Calibrated By Chad Adams Signature  Date: 2/14/2025

APPROVED
By Kayla Myers at 11:11 am, Mar 03, 2025