

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-12202024-004

<p>Gage ID PEA-2905231617 Gage S/N 2905231617 Description 8" Digital Caliper Operating Procedure: OP-110 Unit of Meas. inch Manufacturer insize Cal. Date 12/20/2024 Next Due 6/20/2025 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes</p> <p>Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13607 Job 11812

Standard ID	1.000 "	Uncertainty	Units	Type	V
Limited Use?	No	Minimum	0.999 Nominal	1.000	Maximum 1.001
Ref Type	81PC Gage Blocks	As Found	1.000 Accuracy	0.000	Fail Before No
		As Left	1.000 Accuracy	0.000	Fail After No
Gage ID of Standard	REF-070111				
Std Due Date	1/26/2027	Gage S/N	070111	NIST No.	81004

Standard ID	3.000 "	Uncertainty	Units	Type	V
Limited Use?	No	Minimum	2.999 Nominal	3.000	Maximum 3.001
Ref Type	81PC Gage Blocks	As Found	3.000 Accuracy	0.000	Fail Before No
		As Left	3.000 Accuracy	0.000	Fail After No
Gage ID of Standard	REF-070111				
Std Due Date	1/26/2027	Gage S/N	070111	NIST No.	81004

Standard ID	5.000 "	Uncertainty	Units	Type	V
Limited Use?	No	Minimum	4.999 Nominal	5.000	Maximum 5.001

Ref Type	81PC Gage Blocks	As Found	5.000	Accuracy	0.000	Fail Before	No
		As Left	5.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027 Gage S/N 070111 NIST No. 81004

Standard ID	6.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	5.999	Nominal	6.000	Maximum	6.001
Ref Type	81PC Gage Blocks	As Found	6.000	Accuracy	0.000	Fail Before	No
		As Left	6.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027 Gage S/N 070111 NIST No. 81004


Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	81PC Gage Blocks	As Found	7.000	Accuracy	0.000	Fail Before	No
		As Left	7.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027 Gage S/N 070111 NIST No. 81004

Calibrated By Chad Adams

Signature



Date: 12/20/2024

APPROVED

By Kayla Myers at 2:53 pm, Dec 23, 2024