

**TSP Calibration, Inc.**

1110 S. FM 1788 Suite 38 Midland, TX 79706  
Phone: (432) 296-4992 Email: Chad@tspndt.com

**Certificate of Calibration**

Certificate No TSPW-12202024-001

<b>Gage ID</b> PEA-11212 <b>Gage S/N</b> 11212 <b>Description</b> 50-0-50 gauss meter <b>Operating Procedure:</b> OP-117 <b>Unit of Meas.</b> gauss <b>Manufacturer</b> R.B. Annis <b>Cal. Date</b> 12/20/2024 <b>Next Due</b> 6/20/2025 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b>  <b>Temperature</b> 68 Deg F +/- 2 <b>Humidity</b> 35-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT Solutions
---	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

SO 13607 Job 11812

<b>Standard ID</b> +25 <b>Limited Use?</b> No <b>Ref Type</b> 50-0-50 Gauss Meter	<b>Uncertainty</b> <b>Minimum</b> <b>As Found</b> <b>As Left</b>	<b>Units</b> <b>Nominal</b> <b>Accuracy</b> <b>Accuracy</b>	<b>Type</b> V <b>Maximum</b> 26 <b>Fail Before</b> No <b>Fail After</b> No
<b>Gage ID of Standard</b> REF-TSPSTD-01 <b>Std Due Date</b> 8/8/2025	<b>Gage S/N</b> TSPSTD-01	<b>NIST No.</b> 154783	
<b>Standard ID</b> +50 <b>Limited Use?</b> No <b>Ref Type</b> 50-0-50 Gauss Meter	<b>Uncertainty</b> <b>Minimum</b> <b>As Found</b> <b>As Left</b>	<b>Units</b> <b>Nominal</b> <b>Accuracy</b> <b>Accuracy</b>	<b>Type</b> V <b>Maximum</b> 51 <b>Fail Before</b> No <b>Fail After</b> No
<b>Gage ID of Standard</b> REF-TSPSTD-01 <b>Std Due Date</b> 8/8/2025	<b>Gage S/N</b> TSPSTD-01	<b>NIST No.</b> 154783	
<b>Standard ID</b> -25 <b>Limited Use?</b> No	<b>Uncertainty</b> <b>Minimum</b>	<b>Units</b> <b>Nominal</b>	<b>Type</b> V <b>Maximum</b> 26

Ref Type	50-0-50 Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No

Gage ID of Standard REF-TSPSTD-01

Std Due Date 8/8/2025

Gage S/N TSPSTD-01

NIST No.

154783

Standard ID	-50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	50-0-50 Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No

Gage ID of Standard REF-TSPSTD-01

Std Due Date 8/8/2025


Gage S/N TSPSTD-01

NIST No.

154783

Calibrated By Chad Adams

Signature



Date: 12/20/2024

**APPROVED***By Kayla Myers at 2:51 pm, Dec 23, 2024*