

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-12202024-002

<p>Gage ID PEA-20848 Gage S/N 20848 Description 10 lb Test Weight Operating Procedure: OP-118 Unit of Meas. pounds Manufacturer Parker Cal. Date 12/20/2024 Next Due 12/20/2027 Cal. Freq. 3.00 Years Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13607 Job 11812
AS FOUND: PASSED AS LEFT: PASSED
Using calibrated scale serial no. 8335029565
Test weight as found 10.070 lbs.
Test weight as left 10.070 lbs.
Calibration due date for scale is on 8/8/2025
Scale is traceable to NIST by the following numbers: Doc# 154780

Calibrated By Chad Adams Signature  Date: 12/20/2024

APPROVED
By Kayla Myers at 2:51 pm, Dec 23, 2024